



Certificate of Traceability

Manufacturer Certification

This certifies that *Abbott Diagnostics*, using *HbA1c (WHOLE BLOOD)* *on Architect c8000* has participated in and successfully completed the NGSP certification for manufacturers and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **University of Missouri SRL#9**

The system evaluated was:

Instrument: Architect c8000 analyzer	Calibrator Lot: 57486UQ05, 57486UQ05	Application: %NGSP=HbA1c/Hb*91.48+
Reagent Lot: 57337UQ06, 57337UQ06, 57337UQ06	Calibrator Assigned Values: 5.2%, 11.6%	2.152 WHOLE BLOOD

Date of Certification: May 1, 2021

Certification Expires: May 1, 2022

Handwritten signature of David Sacks.

NGSP Steering Committee Chair

Handwritten signature of Randie R Little PhD.

NGSP Network Coordinator

Handwritten signature of Sharon Amalf.

SRL director/ supervisor